

CERTIFICATE OF COMPLIANCE

Certificate Number 20130722-E347196-E347196
Report Reference E347196-2013-04-19/ A1
Issue Date 2012-September-27

Issued to: STRYKER CORP MEDICAL PRODUCTS DIV
3800 E CENTRE AVE
PORTAGE MI 49002
UNITED STATES

**This is to certify that
representative samples of** COT LOADING DEVICE
MODELS 6390 POWER-LOAD

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10) (Medical
Electrical Equipment – Part 1: General Requirements for
Basic Safety and Essential Performance)


CAN/CSA-C22.2 No. 60601-1 (2008) (Medical Electrical
Equipment – Part 1: General Requirements for Basic Safety
and Essential Performance)

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Classification Mark for the U.S. and Canada should be considered as being covered by UL's Classification and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Classification Mark includes: the UL in a circle symbol:  with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) as indicated in the appropriate UL Directory. The UL Classification Mark for Canada includes: the UL Classification

Mark for Canada:  with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) in English, French, or English/French as indicated in the appropriate UL Directory.



William R. Carney, Director, North American Certification Programs

UL LLC


Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number 20130722-E347196-E347196
Report Reference E347196-2013-04-19/ A1
Issue Date 2012-September-27

Look for the UL Classification Mark on the product. This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.



William R. Carney, Director, North American Certification Programs
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

